

TCT

Certificate of Compliance

Certificate No. : TCT200703E907C

Applicant : Shenzhen Linke Technology Co., Ltd.

Address : A04605, 46F, SaiGe Plaza, Huaqiang North Road, Futian District, Shen Zhen

Manufacturer : Shenzhen Linke Technology Co., Ltd.

Address : A04605, 46F, SaiGe Plaza, Huaqiang North Road, Futian District, Shen Zhen

Product : Fingertip Pulse Oximeter

Model No. : LK-88, LK-89, LK-87

Trade mark : N/A

The above products have been tested by us with listed standards and found in compliance with the council EMC 2014/30/EU. It is possible to use CE marking to demonstrate the compliance with this EMC.

Test standards:	Report(s) Number	Issued By	Issued Date
EN 61326-1:2013 EN 61000-3-2:2014 EN 61000-3-3:2013 EN 61326-2-2:2013	TCT200703E907	TCT	Jul. 07, 2020

The statement is based on a single evaluation of one sample of above mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.



Tomsin

Tomsin/Senior Engineer

Jul. 07, 2020



Shenzhen TCT Testing Technology Co., Ltd.

Hotline: 400-6611-140 Tel: 86-755-27673339 Fax: 86-755-27673332 [Http://www.tct-lab.com](http://www.tct-lab.com)

